




# KALIBRATIERAPPORT

Calibration Report / Kalibrierreport / Rapport d'étalonnage

	<u>Input Standard</u>	<u>Instrument</u>
Tag nr. :		
Manufacturer :	BD/S DM01-A21	ET Ø 100 MM
Type :	Digital ETd 04	7211
Description :	Digital Control	Pressure Gauge
Serial number :	11224402	146952706
Range input :	-1	0
Range output :	5	2,5
Accuracy :	0,05	1,0
Test medium :	Air	
Temperature :	20	
Aantal meetingen :	8	
Certificate nr. :	22-13111-0-1	Measure Temperature 18 – 22°C



Nom. Waarde bar	Input Up Reading	Deviation Up bar	Input Down Reading	Deviation Down bar
0	0,00	0,00	0,00	0,00
0,500	0,501	0,001	0,495	0,005
0,800	0,800	0,000	0,797	0,003
1,000	1,001	0,001	0,996	0,004
1,200	1,199	0,001	1,197	0,003
1,500	1,503	0,003	1,498	0,002
1,700	1,704	0,004	1,702	0,002
2,000	2,006	0,006	2,003	0,003
2,500	2,505	0,005	2,505	0,005

Max Deviation	0,006
Max Allowed Deviation	0,025

Test resultaat: O.K.

Report nr. : 22042801  
 Cal. date : 28-4-2022  
 Customer name :  
 Traceable to : RvA 22-13111-0-1 ISO/IEC 17025

Calibrated by  
 M. J. Biggelaar

Approved by  
 W. Fassbender