

TESTRAPPORT


Rapport nummer

3

	<u>Testinstrument</u>	<u>Instrument</u>
Instr.nummer	2276942	152000346
Fabrikaat	Wika	AB
Omschrijving	Digi Testmanometer	Manometer
Meetbereik	0-2,5 bar	0-2,5 bar
Klasse	0.1	1.0
Klant	TSH Vakopleiding bv	
Getest door	B. Sewkaransing	

Meetpunten	Range Bar	Meetwaarde		Meetfout %	
		Stijgend	Dalend	Stijgend	Dalend
1	0,000	0,000	0,000	0,000	0,000
2	0,500	0,497	0,516	-0,120	0,640
3	1,000	0,996	1,017	-0,160	0,680
4	1,500	1,501	1,520	0,040	0,800
5	2,000	2,012	2,023	0,480	0,920
6	2,500	2,522	2,522	0,880	0,880

Getest volgens DIN en 837-1: 1997



Datum 02-01-25
Ref. klant keuring
Ref. Revin 20240909

REVIN BRIELLE B.V.
't Woud 16
3232 LN Brielle
Tel: 0181-413094
e-mail: info@revinbrielle.com

Revin Brielle B.V. - Industrieterrein 't Woud 16, 3232 LN Brielle
Tel. (0181) 41 30 94 - E-mail info@revinbrielle.com - www.revinsbrielle.com
Bank: Rabo. Rekeningnr. IBAN NR NL45RABO0344868680 - K.v.K. Rotterdam 24237422